


# Replacing a Thermistor in a Core Probe

This document explains the contents of a core probe and how to replace the internals on a Monika Core Probe (4050)



## Replacing the Thermistor

Simply unscrew the choccy block on the side that retains the thermistor, remove the original and replace with the new unit in the original position.

## Routing the Probe Wire

This Diagram shows how to push the probe wire through the central hole on the back plate and stick foam pad to allow the probe wire to be connected to the "choccy block" inside the unit.

